



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Ergonomics

405 Caredean Dr., Horsham, PA 19044

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Dimensional, Electrical, Mass, Force & Weighing, Mechanical, Thermodynamic and Time & Frequency Calibration
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Initial Accreditation Date:

December 22, 2016

Issue Date:

August 01, 2023

Expiration Date:

September 30, 2025

Accreditation No.:

75821

Certificate No.:

L22-578

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjilabs.com



Certificate of Accreditation: Supplement

Ergonomics

405 Caredean Dr., Horsham, PA 19044
 Contact Name: Mr. Rudi Bauknecht Phone: 215-674-0663

Accreditation is granted to the facility to perform the following calibrations:

Dimensional

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Angle Measurements ^F Customer Supplied Product	Up to 90°	0.006 °	Optical Comparator, Angle Blocks Proc #0130, 0189, 0443
Length Measurement – 1D ^F Customer Supplied Product	0.005 mm to 50 mm	11 μ m	Micrometer, Gage Block, Caliper, Optical Comparator, Proc #0130, 0189, 0443
	50 mm to 300 mm	12 μ m	
	300 mm to 1 000 mm	(12 + 4L) μ m	

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure AC Voltage (at the listed frequencies) ^F			Agilent 3458A Procedure 0119, 0150, 0266, 0401
10 Hz to 1 kHz	0.2 V to 750 V	50 mV + 0.4 mV/V	
1 kHz to 20 kHz	0.2 V to 750 V	20 mV + 0.06 mV/V	
20 kHz to 100 kHz	0.2 V to 750 V	0.2 V + 3 mV/V	
100 kHz to 300 kHz	0.2 V to 750 V	0.1 V + 4 mV/V	
300 kHz to 2 MHz	0.2 V to 750 V	0.1 V + 15 mV/V	
Equipment to Measure AC Voltage (at the listed frequencies) ^F			Agilent 3458A, Voltage Divider VD10 dwg 0393, Procedure 0119, 0150, 0266, 0401
10 Hz to 100 kHz	750 V to 10 kV	50 mV + 0.44 mV/V	
Equipment to Measure AC Current (at the listed frequencies) ^F			Agilent 3485A Procedure 0119, 0150, 0266, 0401
10 Hz to 5 kHz ^F	0.02 A to 3 A	6 μ A + 2.1 mA/A	
Equipment to Measure DC Current ^F	0.003 A to 3 A	0.1 mA/A + 1 μ A	
Equipment to Measure Resistance ^F	10 Ω to 10 M Ω	50 $\mu\Omega$ + 8 $\mu\Omega/\Omega$	
Equipment to Measure DC Voltage ^F	0.001 mV to 1 000 V	17 μ V/V	Agilent 3458A, Voltage Divider, Procedure, 0119, 0150, 0266, 0401
	1 000 V to 10 000 V	0.4 μ V/V	
Equipment to Measure Magnetic Field Strength ^F At listed frequency Up to 400 kHz	100 nT to 10 000 nT	0.14 nT + 0.48 % of rdg	Use of a Helmholtz coil Procedure 0129, 0466
	50 V to 5 kV DC	1 V/m + 0.5 % of rdg	Helmholtz coil Reference Plane Procedure 0129, 0466



Certificate of Accreditation: Supplement

Ergonomics

405 Caredean Dr., Horsham, PA 19044

Contact Name: Mr. Rudi Bauknecht Phone: 215-674-0663

Accreditation is granted to the facility to perform the following calibrations:

Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Mass ^F	Up to 300 g	8 mg	Analytic Balance Class II Scale Using ASTM Class I standard Weights Procedure 0139, 0203
	300 g to 3.2 kg	10 mg	
	> 3.2 kg to 10 kg	0.85 g	
	10 kg to 60 kg	1.5 g	

Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Flow Meter of Water ^F	Up to 5 L/min	7.7 mL/min	Analytic Balance Stopwatch, Procedure 0398
	5 L/min to 15 L/min	25 mL/min	
	15 L/min to 50 L/min	77 mL/min	Analytic Balance Stopwatch IEC 60068-2-68 Procedure 01xx
	50 L/min to 110 L/min	300 mL/min	
Force Gauge – Compression/Tension ^F	Up to 2.9 N (0 g to 300 g•f)	0.000 025 N (2.5 mg•f)	ASTM Class 6, Class F Weights, Procedure 0139, 0203
	2.9 N to 11.77 N (300 g to 1.2 kg•f)	0.000 57 N (58 mg•f)	
	11.77 N to 98.067 N (1.2 kg to 10 kg•f)	0.005 7 N (580 mg•f)	
	98.067 N to 222.61 N (10 kg to 60 kg•f)	0.011 N (1.1 g•f)	
Impact Hammer ^F	Up to 2 J (0 Nm to 2 Nm)	0.01 J (0.01 Nm)	HC10 Ljungmann, AS IEC 60068-2-65 Procedure 0128

Thermodynamic

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Thermocouple ^F	961.8 °C	2.1 °C	PTL T 03.89 Certified silver foil- (99.97 % Ag) Procedure 0461

Time & Frequency

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure Frequency ^F	10 Hz to 1 MHz	2.50 Hz + 1% of rdg	Agilent 3458A Procedure 0768



Certificate of Accreditation: Supplement

Ergonomics

405 Caredean Dr., Horsham, PA 19044

Contact Name: Mr. Rudi Bauknecht Phone: 215-674-0663

Accreditation is granted to the facility to perform the following calibrations:

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location. Example: Outside Micrometer^F would mean that the laboratory performs this calibration at its fixed location.
4. The term L represents length in inches or millimeters as appropriate to the uncertainty statement. Also, L

